

AD \_\_\_\_\_

Award Number: DAMD17-02-1-0461

TITLE: A High Resolution Clinical PET with Breast and Whole Body  
Transfigurations

PRINCIPAL INVESTIGATOR: Wai-Hoi Wong, Ph.D

CONTRACTING ORGANIZATION: The University of Texas M.D. Anderson  
Cancer Center  
Houston, TX 77030

REPORT DATE: April 2004

TYPE OF REPORT: Annual

PREPARED FOR: U.S. Army Medical Research and Materiel Command  
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for Public Release;  
Distribution Unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

20040901 051

**REPORT DOCUMENTATION PAGE**Form Approved  
OMB No. 074-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503

**1. AGENCY USE ONLY**  
(Leave blank)**2. REPORT DATE**  
April 2004**3. REPORT TYPE AND DATES COVERED**  
Annual (1 Apr 2003 - 31 Mar 2004)**4. TITLE AND SUBTITLE**

A High Resolution Clinical PET with Breast and Whole Body Transfigurations

**5. FUNDING NUMBERS**

DAMD17-02-1-0461

**6. AUTHOR(S)**

Wai-Hoi Wong, Ph.D.

**7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES)**The University of Texas M.D. Anderson  
Cancer Center  
Houston, TX 77030

E-Mail: gwong@di.mdacc.tmc.edu

**8. PERFORMING ORGANIZATION  
REPORT NUMBER****9. SPONSORING / MONITORING  
AGENCY NAME(S) AND ADDRESS(ES)**U.S. Army Medical Research and Materiel Command  
Fort Detrick, Maryland 21702-5012**10. SPONSORING / MONITORING  
AGENCY REPORT NUMBER****11. SUPPLEMENTARY NOTES**

Original contains color plates: ALL DTIC reproductions will be in black and white

**12a. DISTRIBUTION / AVAILABILITY STATEMENT**

Approved for Public Release; Distribution Unlimited

**12b. DISTRIBUTION CODE****13. ABSTRACT (Maximum 200 Words)**

Despite advances in the last decade, the radiographic diagnosis of breast cancer remains uncertain. Of the annual 600,000 cases referred for biopsy by mammograms each year, 400,000 are unnecessary, costing \$2 billion annually. The diagnosis of breast cancer in young women and women with silicone implants continues to be difficult. Accurate detection of small breast tumors (2-3 mm) is still to be achieved. Positron emission tomography (PET) has the potential to reduce this high healthcare cost, unnecessary painful anxiety, and to improve diagnosis and survivability for women of all ages. We have developed the detector and electronic technology for building an ultrahigh resolution PET camera. We propose to use such technology to construct an ultrahigh resolution PET that has a dedicated breast-diagnosis mode that has 13-26 times higher detection sensitivity than regular PET and an ultrahigh image resolution of 2.5mm compared to the 4.5-6 mm in today's PET cameras. We have already developed a scaled-down engineering prototype PET to confirm the feasibility that 2-3 mm tumor can be detected accurately. We propose to construct a scaled-up clinical version of the design so that it can be used for clinical human trials to confirm the clinical utility.

**14. SUBJECT TERMS**

Breast Cancer

**15. NUMBER OF PAGES**

14

**16. PRICE CODE****17. SECURITY CLASSIFICATION  
OF REPORT**

Unclassified

**18. SECURITY CLASSIFICATION  
OF THIS PAGE**

Unclassified

**19. SECURITY CLASSIFICATION  
OF ABSTRACT**

Unclassified

**20. LIMITATION OF ABSTRACT**

Unlimited

## Table of Contents

<b>Cover.....</b>	<b>1</b>
<b>SF 298.....</b>	<b>2</b>
<b>Table of Contents .....</b>	<b>3</b>
<b>Introduction.....</b>	<b>4</b>
<b>Body.....</b>	<b>4</b>
<b>Key Research Accomplishments.....</b>	<b>9</b>
<b>Reportable Outcomes.....</b>	<b>10</b>
<b>Conclusions.....</b>	<b>12</b>
<b>References.....</b>	<b>13-14</b>
<b>Appendices.....</b>	

## **INTRODUCTION**

Despite advances in the last decade, the radiographic diagnosis of breast cancer remains uncertain. Also, the diagnosis of breast cancer in dense breasts continues to be difficult, whether the density is a result of fibrocystic diseases or young age. Accurate detection of very small breast tumors (2-3 mm) and small metastases, which is essential for survival and breast conservation, is still to be achieved. Positron emission tomography (PET) has the potential to improve in these areas as PET images physiologic differences between tumor and normal tissue, providing > 90% sensitivity and >95% specificity in many recent studies. It also eliminates imaging difficulties for women who have silicone implants and dense breasts, which have negligible effect in PET; hence, PET is useful for high-risk young women and patients with implants.

However, for breast cancer diagnosis, there are technical limitations in current PET cameras that are designed for whole-body tumor staging instead of breast imaging. This goal of this project is to help the development of an ultrahigh resolution convertible-PET design that minimizes the limitations:

- (1) A convertible gantry with a dedicated breast mode that has 10-15 times higher detection sensitivity for breast-tumor activities than a whole-body PET
- (2) An intrinsic image resolution of 2.5-3.0 mm

## **BODY**

The statement of work or task for this period is as follows:


- TASK 1.** Developing a PET camera gantry that can be transformed between the whole-body mode and the breast mode (Months 1-30): Design and Construct the mechanical gantry with a power distribution system (Month 2-24).
- TASK 2.** Fabricating and testing 40,000 small scintillation detectors, and building the detection system (Month 1-24): Produce the scintillation detectors (Month 7-24). Test and quality control for the scintillation detector produced (Month 12-24). Test assembling the detector system with scintillation crystal and photomultipliers integrated together (month 21-24).
- TASK 3.** Developing the high-speed electronic system, 169 circuit boards (Month 1-30): Design, build and test the fast amplifier boards--48 circuit boards (Month 14-18). Design, build and test the HYPER Anger-positioning electronics--48 boards (month 19-24). Design, build and test the programmable coincidence-detection board (Month 12-24)
- TASK 5.** Developing the software and image-process techniques (Month 1-30): Continuously developing the image reconstruction and image processing technique (Month 12-24).

### **FOR TASK-1:**


We have been productive since the funding of this US Army grant that started two year ago. All the tasks scheduled for the period have been accomplished. Much progress has been made in developing this ultrahigh resolution transformable PET camera. In this one year, our

research effort have resulted in the submission of peer-reviewed scientific papers; 11 of these 17 papers is already in print or accepted to be published, with the other 6 still in the peer-review process.


The transformable rotating PET gantry design is finished [1,2]. The 3-D mechanical design drawings are shown in fig.1-4. These figures are actual 3-D mechanical drawings ("blue print") with correct dimensions and not just drawings.





**FIG. 1** The PET in whole-body imaging mode



**FIG. 2** The PET in breast imaging mode




**FIG. 3** Back of the rotating gantry carrying the electronics




**FIG. 4** This shows how two neighboring rectangular detector modules are rotated and meshed into each other so that the same detector modules can create a big detector ring for wholebody imaging and a small ring for breast imaging.

The detail design of the detector module is shown in fig.5 [1,2]. Two such test-modules were constructed with all crystals, photomultipliers, front-end fast amplifier and control electronics incorporated (fig.6). The 2 test modules were mated together accurately mechanically with no gaps between the two detector modules to facilitate the body-breast conversion (fig.7).




**FIG. 5** Internal design of module




**FIG. 6** Module construction

#### **FOR TASK-2:**

All the position-sensitive BGO detector blocks with a total of 40,000 individual crystals have been constructed [3-4]. Each detector module carries 66 detector blocks and each block has up to 64 crystals. Construction of the position-sensitive detectors has been the most labor intensive part of the project. A constructed detector bank, 66 blocks 3168 crystals, is shown in fig.8 [3-4].




**FIG. 7** Mechanical mating of two test-modules



**FIG. 8** A detector bank with 3168 position-sensitive detectors

The detector-position decoding of the constructed detector blocks were tested for the 40,000 position-sensitive crystals [3] using the prototype electronics we designed and constructed for this PET camera. The crystal-position decoding test results have been excellent (fig.9).




**FIG. 9** Crystal-position decoding results

### **FOR TASK-3:**

#### **Electronic Design and Construction**

The patented high-speed pileup-recovery front-end electronics and decoding-electronics (HYPER) have been designed and all 12 sets of HYPER have been constructed (fig.10) [5-8]. This front-end electronics were tested and they performed well and were used to obtain the results shown in fig. 9. The coincidence-timing and data sorting electronics has also been developed and built for this transformable PET (fig.11) [9].




**FIG. 10** 3 of the 12 sets of HYPER electronics.



**FIG. 11** Coincidence electronics

We have done partial testing on it. The full test cannot be performed until the entire system are put together. The partial test on processing dead-time and event-sorting throughput showed that the circuit can process a maximum of 36 million single-events per second hitting the whole detector system and process 5 million coincidence-event per second. Both of these figures are much higher than most clinical human studies.

We also developed a high-speed system-wide detector-tuning system for this PET camera and this BGO-detector-tuning system should take 1-2 minutes to retune all 40,000 detectors instead of the 4-6 hours in current commercial systems, thereby allowing the whole PET to be recalibrated for every patient to insure optimal imaging quality for every patient, instead of 1-3 months in current practice. This tuning system (using internal LED light) has been developed and tested and recently submitted for journal publication. Test results showed that all 40,000 detectors in the camera can be recalibrated in 1 minute [10-11]. The detector-position-decoding test "before and after tuning" was excellent (fig. 12).



**FIG. 12** Before tuning, crystal positions were moved to the right and caused position-decoding errors. After tuning, the crystals positions were restored to the original positions

Because of the ultrahigh image resolution of the system, we are developing a breathing-cycle gating system to correct for the patient breathing moving. This system used a sensor in the patient's nostril air flow and gating electronics to correct for the patient breathing movement (fig.13). A paper has been submitted for journal publication [12-13]. Test results of the system, using our existing prototype PET, showed that this gating-system increased lesion uptake in the image by an average of 2.2 time for 3-5 mm lesions.




FIG. 13 LEFT---Nostril-air breathing cycle gating system improved lesion uptake in image

### **FOR TASK-5:**

We continue the development of the image reconstruction and image processing. We have developed the software and methodology to study lesion detection in ultrahigh resolution PET cameras. We used this software to evaluate ways to achieve better lesion detection and published one peer-reviewed paper [14]. Four different image-reconstruction software and different image processing filters were implemented, tested and compared. For this research effort in this period, 2 abstracts were presented in conferences [15-16] and another peer-reviewed journal publication [17] was submitted.

### **KEY RESEARCH ACCOMPLISHMENTS**

- The very complex transformable-PET mechanical design has been accomplished and 2 of the 12 detector heads (module) have been fabricated and put together and are being debugged and tested. The test results will be fed back for improving the construction of the rest of the 10 detector heads (both electronics and mechanical).
- All the 40,000 position-sensitive detectors have been constructed and tested. The results have been satisfactory.
- The very complex front-end electronics (HYPER) has been developed and fabricated. The programmable coincidence-detection electronics has also been developed and fabricated. These electronics are currently being tested with the 2 prototype detector head.
- A new detector tuning system has been developed. This system can tune all the detectors in the system in less than 1 minute, which is significant improvement to exisating commercial systems that take many hours to tune a PET with far fewer detectors
- A breathing-cycle movement-compensation method has also been developed to improve lesion detections in our system

- Four different types of image processing software have been implemented for the camera under construction.

## **REPORTABLE OUTCOMES**

### **Peer-Reviewed Journal Publications (8 in print and 7 submit):**

Uribe, J., Aykac, M., Baghaei, H., Li, H., Wang, Y., Liu, Y., Wong, V., Xing, T., Ramirez, R., Wong, W-H. Inexpensive Position Sensitive Detector Block for Dedicated PET Cameras Using 40 mm Diameter PMT in Quadrant Sharing Configuration. *IEEE Transactions on Nuclear Science*, 50(3), 367-372, June, 2003.

Li, H., Xing, T., Liu, Y., Wang, Y., Baghaei, H., Uribe, J., Wong, W-H. A Hotlink/Network PC Data Acquisition and Image Reconstruction System for a High Resolution Whole-Body PET with Respiratory and ECG-Gated Performance. *IEEE Transactions on Nuclear Science*, 50(3), 393-397, June, 2003.

Uribe, J., Li, H., Xing, T., Liu, Y., Baghaei, H., Wang, Y., Aykac, M., Ramirez, R., Wong, W-H. Signal Characteristics of Individual Crystals in a High Resolution BGO Detector Design Using PMT-Quadrant Sharing. *IEEE Transactions on Nuclear Science*, 50(3), 355-361, June, 2003.

Li, H., Wong, W-H., Wang, Y., Liu, Y., Xing, T., Uribe, J., Baghaei, H. Front-End Electronic Design Based on High-Yield-Pileup-Event-Recovery Method for a High-Resolution PET Camera with PMT-Quadrant-Sharing Detector Modules. Submitted for publication to *IEEE Transactions on Nuclear Science*.

Baghaei, H., Uribe, J., Li, H., Aykac, M., Wang, Y., Liu, Y., Xing, T., Wong, W-H. Effects of Attenuation Correction and 3D-Reconstruction Algorithms on Brain Lesions Detectability. Submitted for publication to *IEEE Transactions on Nuclear Science*.

Liu, Y., Li, H., Wang, Y., Xing, T., Uribe, J., Baghaei, H., Wong, W-H. A Programmable High-Resolution Ultra Fast Delay Generator. *IEEE Transactions on Nuclear Science*, 50(5), 1487-1490, October, 2003.

Baghaei, H., Wong, W-H., Aykac, M., Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T. Brain Lesions Detectability Studies with a High Resolution PET Operating in No-Septa and Partial-Septa Configurations. *IEEE Transactions on Nuclear Science*, 50(5), 1364-1369, October, 2003.

Uribe, J., Wong, W-H., Baghaei, H., Farrell, R., Li, H., Aykac, M., Bilgen, D., Liu, Y., Wang, Y., Xing, T. An Efficient Detector Production Method For Position-Sensitive Scintillation Detector Arrays with 98% Detector Packing Fraction. *IEEE Transactions on Nuclear Science*, 50(5), 1469-1476, October, 2003.

Wang, Y., Li, H., Liu, Y., Xing, T., Uribe, J., Baghaei, H., Wong, W-H. A Modular Low Dead-Time Coincidence System for High Resolution PET Cameras. IEEE Transactions on Nuclear Science, 50(5), 1386-1391, October, 2003.

Aykac, M., Li, H., Uribe, J., Wang, Y., Baghaei, H., Liu, Y., Xing, T., Wong, W-H. A Study of Coincidence Line Spread Function (CLSF) Estimation for Small Scintillators Using Quadrant Sharing Technique. IEEE Transactions on Nuclear Science, 50(5), 1331-1338, October, 2003.

Uribe, J., Xie, S., Li, H., Baghaei, H., Wang, Y., Liu, Y., Xing, T., Ramirez, R., Wong, W-H. Gantry Design with Accurate Crystal Positioning for a High-Resolution Transformable PET Camera. Submitted for publication to IEEE Transactions on Nuclear Science.

Baghaei, H., Wong, W-H, Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Ramirez, R., Xie, S., Kim, S. A Comparison of Four Image Reconstruction Algorithms for 3-D PET Imaging Using Phantom Data. Submitted for publication to IEEE Transactions on Nuclear Science.

Li, H., Liu, Y., Xing, T., Wang, Y., Uribe, J., Baghaei, H., Xie, S., Ramirez, R., Wong, W-H. An Instantaneous Photomultiplier Gain Calibration Method for PET or Gamma Camera Detectors Using a LED Network. Submitted for publication to IEEE Transactions on Nuclear Science.

Liu, Y., Li, H., Wang, Y., Xing, T., Xie, S., Uribe, J., Baghaei, H., Ramirez, R., Kim, S., Wong, W-H. A Gain-Programmable Transit-Time-Stable and Temperature-Stable PMT Voltage Divider. Submitted for publication to IEEE Transactions on Nuclear Science.

Wang, Y., Baghaei, H., Li, H., Liu, Y., Xing, T., Uribe, J., Ramirez, R., Xie, S., Kim S., Wong, W-H. A Simple Respiration Gating Technique and its Application in High-Resolution PET Camera. Submitted for publication to IEEE Transactions on Nuclear Science.

#### **Abstract Presented In Conferences (9 total):**

Wang, Y., Wong, W-H., Uribe, J., Li, H., Baghaei, H., Liu, Y., Xing, T., Farrell, R., Xie, S. Energy and Spatial Resolution Relationship Comparison Study on BGO, GSO and LSO. Society of Nuclear Medicine 50th Annual Meeting, New Orleans, LA. Abstract Book, vol. 44, no. 5 (supplement), p. 283P, June, 2003.

Baghaei, H., Wong, W-H., Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Farrell, R., Xie, S. A Comparison of BGO, GSO and LSO Crystals in Photopeak Efficiency and Spatial Resolution for a High Resolution PET Scanner Using Monte Carlo Simulation. Society of Nuclear Medicine 50th Annual Meeting, New Orleans, LA. Abstract Book, vol. 44, no. 5 (supplement), p. 284P, June, 2003.

Baghaei, H., Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Farrell, R., Xie, S., Wong, W-H. A Comparison of the 3D-OSEM, 3DRP and Rebinning Algorithms for Brain Lesion Detectability.

Society of Nuclear Medicine 50th Annual Meeting, New Orleans, LA. Abstract Book, vol. 44, no. 5 (supplement), p. 278P, June, 2003.

Li, H., Baghaei, H., Liu, Y., Xing, T., Wang, Y., Uribe, J., Xie, S., Farrell, R., Wong, W-H. A Parallel Data Acquisition and Image Reconstruction System for PET Cameras. Society of Nuclear Medicine 50th Annual Meeting, New Orleans, LA. Abstract Book, vol. 44, no. 5 (supplement), p. 283P, June, 2003.

Uribe, J., Xie, S., Li, H., Baghaei, H., Wang, Y., Liu, Y., Xing, T., Ramirez, R., Bushman, M., Kim, S., Wong, W-H. Gantry Design with Accurate Crystal Positioning for a High-Resolution Transformable PET Camera. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.

Baghaei, H., Wong, W-H, Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Farrell, R., Xie, S. A Comparison of Four Image Reconstruction Algorithms for Detection of Small Lesions in Brain Phantoms. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.

Li, H., Liu, Y., Xing, T., Wang, Y., Uribe, J., Baghaei, H., Xie, S., Ramirez, R., Wong, W-H. An Instantaneous Photomultiplier Gain Calibration Method for PET or Gamma Camera Detectors Using a LED Network. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.

Liu, Y., Li, H., Wang, Y., Xing, T., Xie, S., Uribe, J., Baghaei, H., Ramirez, R., Wong, W-H. A Gain-Programmable Transit-Time-Stable and Temperature-Stable PMT Voltage Divider. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.

Wang, Y., Baghaei, H., Li, H., Liu, Y., Xing, T., Uribe, J., Farrell, R., Xie, S., Wong, W-H. A Simple Respiration Gating Technique and its Application in High-Resolution PET Camera. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.

## **CONCLUSIONS**

We have accomplished more than the scheduled tasks planned. The research led to 9 abstract presentations in scientific conferences, 7 papers submitted for peer-reviewed publications and 8 papers published in peer-reviewed journals. The very complex transformable-PET mechanical design has been accomplished, 2 of the 12 detector heads (module) have been fabricated and put together and are being debugged and tested. All the 40,000 position-sensitive detectors have been constructed and tested. The very complex front-end electronics (HYPER) has been developed and fabricated. The programmable coincidence-detection electronics has also been developed and fabricated. A new detector tuning system has been developed. This system can tune all the detectors in the system in less than 1 minute, which is significant improvement to existing commercial systems that take many hours to tune a PET with far fewer detectors. This auto-tuning invention can also be adopted by current commercial systems and allows the whole

scanner to be retuned for every patient to assure that the scanner is in its optimal operating condition providing the highest quality images. Currently, commercial scanners are tuned once every few weeks to once every few months because of the long time required for tuning. A breathing-cycle movement-compensation method has also been developed to improve lesion detections in our system, which minimizes image blurring and artifacts created by patient movement during a scan. This methodology can also be used in a regular PET camera for improving cancer detectability. We have accomplished about 60% of the work needed to finish the proposed ultrahigh resolution transformable PET camera that can be transformed into a dedicated breast-PET camera with 15 times higher detection sensitivity and with 2-3 mm image resolution compared to the 5-6mm resolution of current commercial PET. This can lead to earlier and more accurate detection of breast cancer, which improves survivability.

## **REFERENCES**

- [1] Uribe, J., Xie, S., Li, H., Baghaei, H., Wang, Y., Liu, Y., Xing, T., Ramirez, R., Bushman, M., Kim, S., Wong, W-H. Gantry Design with Accurate Crystal Positioning for a High-Resolution Transformable PET Camera. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.
- [2] Uribe, J., Xie, S., Li, H., Baghaei, H., Wang, Y., Liu, Y., Xing, T., Ramirez, R., Wong, W-H. Gantry Design with Accurate Crystal Positioning for a High-Resolution Transformable PET Camera. Submitted for publication to IEEE Transactions on Nuclear Science.
- [3] Uribe, J., Li, H., Xing, T., Liu, Y., Baghaei, H., Wang, Y., Aykac, M., Ramirez, R., Wong, W-H. Signal Characteristics of Individual Crystals in a High Resolution BGO Detector Design Using PMT-Quadrant Sharing. IEEE Transactions on Nuclear Science, 50(3), 355-361, June, 2003.
- [4] Uribe, J., Wong, W-H., Baghaei, H., Farrell, R., Li, H., Aykac, M., Bilgen, D., Liu, Y., Wang, Y., Xing, T. An Efficient Detector Production Method For Position-Sensitive Scintillation Detector Arrays with 98% Detector Packing Fraction. IEEE Transactions on Nuclear Science, 50(5), 1469-1476, October, 2003.
- [5] Li, H., Xing, T., Liu, Y., Wang, Y., Baghaei, H., Uribe, J., Wong, W-H. A Hotlink/Network PC Data Acquisition and Image Reconstruction System for a High Resolution Whole-Body PET with Respiratory and ECG-Gated Performance. IEEE Transactions on Nuclear Science, 50(3), 393-397, June, 2003.
- [6] Li, H., Wong, W-H., Wang, Y., Liu, Y., Xing, T., Uribe, J., Baghaei, H. Front-End Electronic Design Based on High-Yield-Pileup-Event-Recovery Method for a High-Resolution PET Camera with PMT-Quadrant-Sharing Detector Modules. Submitted for publication to IEEE Transactions on Nuclear Science.
- [7] Liu, Y., Li, H., Wang, Y., Xing, T., Uribe, J., Baghaei, H., Wong, W-H. A Programmable High-Resolution Ultra Fast Delay Generator. IEEE Transactions on Nuclear Science, 50(5), 1487-1490, October, 2003.

- [8] Liu, Y., Li, H., Wang, Y., Xing, T., Xie, S., Uribe, J., Baghaei, H., Ramirez, R., Wong, W-H. A Gain-Programmable Transit-Time-Stable and Temperature-Stable PMT Voltage Divider. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.
- [9] Wang, Y., Li, H., Liu, Y., Xing, T., Uribe, J., Baghaei, H., Wong, W-H. A Modular Low Dead-Time Coincidence System for High Resolution PET Cameras. IEEE Transactions on Nuclear Science, 50(5), 1386-1391, October, 2003.
- [10] Li, H., Liu, Y., Xing, T., Wang, Y., Uribe, J., Baghaei, H., Xie, S., Ramirez, R., Wong, W-H. An Instantaneous Photomultiplier Gain Calibration Method for PET or Gamma Camera Detectors Using a LED Network. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.
- [11] Li, H., Liu, Y., Xing, T., Wang, Y., Uribe, J., Baghaei, H., Xie, S., Ramirez, R., Wong, W-H. An Instantaneous Photomultiplier Gain Calibration Method for PET or Gamma Camera Detectors Using a LED Network. Submitted for publication to IEEE Transactions on Nuclear Science.
- [12] Wang, Y., Baghaei, H., Li, H., Liu, Y., Xing, T., Uribe, J., Farrell, R., Xie, S., Wong, W-H. A Simple Respiration Gating Technique and its Application in High-Resolution PET Camera. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.
- [13] Wang, Y., Baghaei, H., Li, H., Liu, Y., Xing, T., Uribe, J., Ramirez, R., Xie, S., Kim S., Wong, W-H. A Simple Respiration Gating Technique and its Application in High Resolution PET Camera. Submitted for publication to IEEE Transactions on Nuclear Science.
- [14] Baghaei, H., Wong, W-H., Aykac, M., Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T. Brain Lesions Detectability Studies with a High Resolution PET Operating in No-Septa and Partial-Septa Configurations. IEEE Transactions on Nuclear Science, 50(5), 1364-1369, October, 2003.
- [15] Baghaei, H., Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Farrell, R., Xie, S., Wong, W-H. A Comparison of the 3D-OSEM, 3DRP and Rebinning Algorithms for Brain Lesion Detectability. Society of Nuclear Medicine 50th Annual Meeting, New Orleans, LA. Abstract Book, vol. 44, no. 5 (supplement), p. 278P, June, 2003.
- [16] Baghaei, H., Wong, W-H, Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Farrell, R., Xie, S. A Comparison of Four Image Reconstruction Algorithms for Detection of Small Lesions in Brain Phantoms. IEEE Nuclear Science Symposium and Medical Imaging conference record, vol. pp. - , Portland. OR, October 19-25, 2003.
- [17] Baghaei, H., Wong, W-H, Uribe, J., Li, H., Wang, Y., Liu, Y., Xing, T., Ramirez, R., Xie, S., Kim, S. A Comparison of Four Image Reconstruction Algorithms for 3-D PET Imaging Using Phantom Data. Submitted for publication to IEEE Transactions on Nuclear Science.